

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00543	Calib. Date: 03/23/2015
Certification File No.:	00544	Calib. No.: 00025
Linearity File No.:	00545	Cert. Date: 03/23/2015
Solution File No.:	00562	Cert. No.: 00017
Sequential File No.:	00562	Lin. Date: 03/23/2015
		Lin. No.: 00021
		Soln. Date: 07/22/2015
		Soln. No.: 00127
		File Date: 07/22/2015

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXC S3-0164
Control Solution %: 0.100%		Expires: 11/12/2015
Solution Control Lot: 13K125		Bottle No.: 1240

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:47D		
Control 1 EC	0.100%	00:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	00:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:48D		
Control 2 EC	0.100%	00:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:49D		
Control 3 EC	0.101%	00:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	00:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:50D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: 

Badge No.: 96

Date: 07/22/2015

